



Appendix A

LAST NAME ZERO
FIRST NAME TEST

FIT TEST REPORT

Fit Test Information

ID NUMBER	1234	CUSTOM1	
LAST NAME	ZERO	CUSTOM2	
FIRST NAME	TEST	CUSTOM3	
COMPANY		CUSTOM4	
LOCATION			
NOTE	CONFIRMS MAX FIT FACTOR		
TEST DATE	02/15/2000	PORTACOUNT S/N	41834
TEST TIME	09:24	NS5 COMPANION	N
DUE DATE	02/15/2001		
RESPIRATOR		PROTOCOL	DEFAULT
MANUFACTURER	TSI	PASS LEVEL	29CFR1910.134
MODEL	ZERO FILTER		500
MASK STYLE	1602066		
MASK SIZE	FILTER		
APPROVAL			
EFF. <99%	N		

<u>EXERCISE</u>	<u>DURATION (SEC)</u>	<u>FIT FACTOR</u>	<u>PASS</u>
NORMAL BREATHING	60	557000	Y
DEEP BREATHING	60	557000	Y
READING OR TALKING	60	564000	Y
HEAD UP AND DOWN	60	565000	Y
HEAD SIDE TO SIDE	60	560000	Y
GIMMACE	30	141000	
BEND	60	563000	Y
NORMAL BREATHING	60	561000	Y
OVERALL FF		409000	Y

FIT TEST OPERATOR WEINSTEIN DATE _____

NAME TEST ZERO DATE _____